

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 08/22/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:55
INTERNAL STANDARD	VERIFIED	10:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:55
BLANK TEST	0.00	10:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:57
BLANK TEST	0.00	10:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:58
BLANK TEST	0.00	10:59

