

SOLUTION CHANGE

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 08/31/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:20
INTERNAL STANDARD	VERIFIED	17:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:21
BLANK TEST	0.00	17:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:22
BLANK TEST	0.00	17:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:23
BLANK TEST	0.00	17:24

