

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 09/05/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:24
INTERNAL STANDARD	VERIFIED	10:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:24
BLANK TEST	0.00	10:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:25
BLANK TEST	0.00	10:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:26
BLANK TEST	0.00	10:27

