

SUPERVISOR TEST

LAKE GEORGE
PARK COMMISSION
DATAMASTER dmt: 300391

Date: 09/26/2022

OPERATOR NAME:
SHANE L ROSS
Permit #: 12663
Operator Agency/Dept: LAKE GEORGE PK COMM
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:15
INTERNAL STANDARD	VERIFIED	11:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:16
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:17
BLANK TEST	0.00	11:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:18
BLANK TEST	0.00	11:19

