

SOLUTION CHANGE

LAKE GEORGE  
PARK COMMISSION  
DATAMASTER dmt: 300391

Date: 10/02/2022

OPERATOR NAME:  
SHANE L ROSS  
Permit #: 12663  
Operator Agency/Dept: LAKE GEORGE PK COMM  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:16
INTERNAL STANDARD	VERIFIED	13:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:16
BLANK TEST	0.00	13:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:18
BLANK TEST	0.00	13:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:19
BLANK TEST	0.00	13:20

