

SUPERVISOR TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391

Date: 10/10/2022

OPERATOR NAME:
 SHANE L ROSS
 Permit #: 12663
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:16
INTERNAL STANDARD	VERIFIED	09:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:17
BLANK TEST	0.00	09:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:19
BLANK TEST	0.00	09:20

