

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: WPC

Instrument Serial #: 300391

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
7/22	22100	34.0	34.01	34.01	MP5081	.10	.10	.10	PACQUÉ	In Service YES
8/1	22190	34.01	34.01	34.01	MP5081	.10	.10	.10	ROSS	YES
8/8	22190	34.01	34.01	34.01	MP5081	.09	.09	.09	ROSS	
8/15	22190	34.01	34.01	34.01	MP5081	.10	.10	.10	ROSS	
8/22	22190	34.01	34.01	34.01	MP5081	.10	.10	.10	ROSS	
8/29	22190	34.01	34.01	34.01	MP5081	.09	.09	.10	ROSS	YES
9/5	22190	34.01	34.01	34.01	MP5081	.10	.09	.10	ROSS	
9/12	22190	34.01	34.01	34.01	MP5081	.10	.10	.10	ROSS	
9/19	22190	34.01	34.01	34.01	MP5081	.10	.09	.10	ROSS	
9/26	22190	34.01	34.01	34.01	MP5081	.09	.09	.09	ROSS	
10/2	22260	34.01	34.01	34.01	MP5081	.10	.10	.10	ROSS	YES
10/10	22260	34.01	34.01	34.01	MP5081	.10	.10	.10	ROSS	

* - Please enter a √ to denote simulator solution change taking place

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: LOPL

Instrument Serial #: 300391

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
5/14/22	22030	34.00	34.00	34.00	MP5081	.10	.10	.10	Ross	IN SERVICE
5/17/22	22030	34.00	34.00	34.00	MP5081	.10	.10	.10	Ross	
5/16/22	22030	34.01	34.01	34.00	MP5081	.10	.10	.10	Ross	
5/24/22	22030	34.01	34.00	34.01	MP5081	.10	.10	.10	Ross	
5/30/22	22030	34.01	34.01	34.01	MP5081	.09	.10	.10	Ross	
6/1/22	22030	34.0	34.00	34.00	MP5081	.10	.10	.10	Ross	YES
6/7/22	22100	34.01	34.01	34.01	MP5081	.10	.10	.10	Ross	YES
6/13/22	22100	34.01	34.00	34.01	MP5081	.10	.10	.10	Ross	
6/20/22	22100	34.01	34.01	34.01	MP5081	.10	.10	.10	Ross	
6/27/22	22100	34.01	34.01	34.01	MP5081	.10	.10	.10	Ross	
7/3/22	22100	34.01	34.01	34.01	MP5081	.09	.10	.10	Ross	YES
7/11/22	22100	34.01	34.00	34.01	MP5081	.09	.10	.10	Ross	

* - Please enter a √ to denote simulator solution change taking place

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: L6PC

Instrument Serial #: 300391

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
10/19	22260	34.01	34.01	34.00	MP5081	.09	.09	.10	Ross	

* - Please enter a √ to denote simulator solution change taking place