



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 300391, performed on April 24, 2020 by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under
ESRA by Joshua Vinehout
on 4/29/2020 8:23 AM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission

Manufacturer: NAT'L PATENT ANALYTICAL SYS

Model: DATAMASTER DMT

Date: April 24, 2020

Serial Number: 300391

Reference Solution Lot Number: 19281

Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	34.01 degrees centigrade
2	0.100 %	34.01 degrees centigrade
3	0.100 %	34.01 degrees centigrade
4	0.100 %	34.01 degrees centigrade

Reference Solution Lot Number: 20020

Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.080 %	34.01 degrees centigrade
2	0.079 %	34.01 degrees centigrade
3	0.079 %	34.01 degrees centigrade
4	0.080 %	34.01 degrees centigrade

Reference Solution Lot Number: 19210

Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.177 %	34.01 degrees centigrade
2	0.178 %	34.01 degrees centigrade
3	0.178 %	34.01 degrees centigrade
4	0.178 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 300391, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under
ESRA by Martin Schilling
on 4/24/2020 1:27 PM

Martin Schilling
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45926

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 4/28/2020 11:05 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 300391
Date: April 24, 2020

1. General Appearance and condition of this device -----: OK

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: OK

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION



Performed by:

Digitally signed under
ESRA by Martin Schilling
on 4/24/2020 1:27 PM

Martin Schilling
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45926

Reviewed by:

Digitally signed under
ESRA by Frank Sgarlata
on 4/28/2020 11:05 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391





Date: 04/24/2020

OPERATOR NAME:
 MARTIN J SCHILLING
 Permit #: 45926
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 20020

Nominal = 0.080 g/210L

BLANK TEST	0.000	09:03
INTERNAL STANDARD	VERIFIED	09:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	09:03
BLANK TEST	0.000	09:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	09:04
BLANK TEST	0.000	09:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	09:05
BLANK TEST	0.000	09:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	09:07
BLANK TEST	0.000	09:07

Average = 0.0795
 Std Dev = 0.0006



TECHNICIAN TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391





Date: 04/24/2020

OPERATOR NAME:
 MARTIN J SCHILLING
 Permit #: 45926
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 19281

Nominal = 0.100 g/210L

BLANK TEST	0.000	08:55
INTERNAL STANDARD	VERIFIED	08:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.100	08:56
BLANK TEST	0.000	08:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.100	08:57
BLANK TEST	0.000	08:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.100	08:58
BLANK TEST	0.000	08:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.100	08:59
BLANK TEST	0.000	09:00

Average = 0.1000
 Std Dev = 0.0000



TECHNICIAN TEST

LAKE GEORGE
 PARK COMMISSION
 DATAMASTER dmt: 300391

Date: 04/24/2020

OPERATOR NAME:
 MARTIN J SCHILLING
 Permit #: 45926
 Operator Agency/Dept: LAKE GEORGE PK COMM
 Reference Standard #: 19210

Nominal = 0.180 g/210L

BLANK TEST	0.000	09:12
INTERNAL STANDARD	VERIFIED	09:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.177	09:12
BLANK TEST	0.000	09:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.178	09:14
BLANK TEST	0.000	09:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.178	09:15
BLANK TEST	0.000	09:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.178	09:16
BLANK TEST	0.000	09:17

Average = 0.1778
 Std Dev = 0.0005

