



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR FST SCREENER, serial number 043713, performed on March 13, 2023 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 03/22/2023 01:32 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission
Manufacturer: INTOXIMETERS INC.
Model: ALCO-SENSOR FST
Serial Number: 043713
Date: March 13, 2023

Reference Solution Lot Number : 22390
Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	34.00 degrees centigrade
2	0.099 %	34.00 degrees centigrade
3	0.099 %	34.00 degrees centigrade
4	0.096 %	34.00 degrees centigrade



**Division of Criminal
Justice Services**

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR FST, serial number 043713, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.

A handwritten signature in black ink that reads "Joe Iavarone".

Digitally signed under ESRA
by Joseph Iavarone
on 03/13/2023 12:30 PM

**Joseph Iavarone
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 043713

Date: March 13, 2023

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : RE-ADJUSTED,
VERIFIED CALIBRATION**



**Division of Criminal
Justice Services**

A handwritten signature in cursive script that reads "Joe Iavarone".

Digitally signed under ESRA
by Joseph Iavarone
on 03/13/2023 12:30 PM

**Joseph Iavarone
Highway Safety Equipment Technician
Office of Public Safety**

