



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, **Steven J. Carluccio**, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of **GUTH LABORATORIES, 12V500 SIMULATOR**, serial number **MP5081**, performed on **May 02, 2023** by Office of Public Safety employee **Martin Schilling**, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 05/11/2023 10:19 AM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission
Manufacturer: GUTH LABORATORIES
Model: 12V500
Serial Number: MP5081
Date: May 02, 2023

Breath Simulator Temperature
34.02 degrees centigrade



**Division of Criminal
Justice Services**

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP5081, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.

A handwritten signature in black ink, appearing to read "Martin Schilling".

Digitally signed under ESRA
by Martin Schilling
on 05/02/2023 12:45 PM

Martin Schilling
Highway Safety Equipment Technician
Office of Public Safety

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP5081

Date: May 02, 2023

1. General appearance and condition of this device -----: OK

2. Check propeller rotation -----: OK

3. Check for air leaks -----: OK

4. Check for proper temperature -----: OK

**5. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**



**Division of Criminal
Justice Services**

A handwritten signature in black ink that reads "Martin Schilling".

Digitally signed under ESRA
by Martin Schilling
on 05/02/2023 12:45 PM

**Martin Schilling
Highway Safety Equipment Technician
Office of Public Safety**

625 East Bunker Court
 Vernon Hills, Illinois 60061
 PH: 866-466-6225
 Fax: 847-327-2993
 www.innocalsolutions.com

NIST Traceable Calibration Report

REPORT NUMBER
1700073

Reference Number: 115843400
 PO Number:

NY State Div Of Criminal Justice Svcs
 80 South Swan Street
 Albany, NY 12210 United States

Manufacturer:	Alpha Technics	Calibration Date	06/17/2022
Model Number:	91-5443	Calibration Due Date	06/17/2023
Description:	Temperature, Thermometer	Condition As Found:	In Tolerance
Asset Number:	CP318227	Condition As Left:	In Tolerance, No adjustment
Serial Number:	20BMM01710		
Procedure:	DS Alpha Technics 5000 Series f/NY State DCJS		
Remarks:			

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP50219	Hart Scientific	1590	Temperature, Super Thermometer	3/10/2021	3/31/2023
CP32036	Hart Scientific	5628	Temperature, Probe	3/03/2021	10/31/2022
CP106957	Hart Scientific	5699	Temperature, Probe, SPRT	3/19/2021	4/30/2023
CP32033	Hart Scientific	5628	Temperature, Probe	5/24/2021	3/31/2023
CP32034	Hart Scientific	5628	Temperature, Probe	11/16/2021	5/31/2023

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Temperature	-15.000 °C					
As Found & As Left	-15.000	-15.00		-15.050 to -14.950 °C	4.3:1	± 0.012 °C
	0.000 °C					
As Found & As Left	0.000	-0.00		-0.030 to 0.030 °C	4.1:1	± 0.0074 °C
	75.000 °C					
As Found & As Left	75.000	74.99		74.970 to 75.030 °C	2.7:1	± 0.011 °C
	125.000 °C					
As Found & As Left	125.000	124.99		124.950 to 125.050 °C	3.9:1	± 0.013 °C

Temperature: 23 °C
 Humidity: 55 %RH
 Rpt. No.: 1700073

Calibration Performed By:				Quality Reviewer:	
Jennifer Sittinger		Metrologist		Szplit, Tony	6/17/2022
Name	ID #	Title	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, ISO 9002-94, and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

*MS
5/2/23*

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency Lake George Park Commission		Today's Date 5.2.23
Address 75 Fort George Rd.		City, State, ZIP Lake George, NY 12845
Contact Person		Contact Telephone 518-668-9347
Contact Email		Best Way to Reach You During Appointment Period

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number 12V500		Instrument Serial Number MP5081
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair
Brief Description of Malfunction		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT SIM Tubes Keyboard Other:		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:
<i>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</i>		
Signature NA Re-Issue for Season		Date NA 5.2.23

DCJS USE ONLY

Export °C <u>34.01</u> <u>34.01</u> <u>34.01</u>	Digital °C <u>34.01</u> <u>34.02</u> <u>34.02</u>	Thermometer: <u>20BMM01710</u> Calibration Due: <u>6.17-23</u> Air Leak Test: <u>OK</u> Propeller Rotation: <u>OK</u>
<i>Opening cert for season</i>		