



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of GUTH LABORATORIES, 12V500 SIMULATOR, serial number MP5081, performed on October 24, 2023 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 11/08/2023 09:19 AM

**Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission
Manufacturer: GUTH LABORATORIES
Model: 12V500
Serial Number: MP5081
Date: October 24, 2023

Breath Simulator Temperature
34.03 degrees centigrade



**Division of Criminal
Justice Services**

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of GUTH LABORATORIES simulator model 12V500, serial number MP5081, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified simulator was performed, or within a reasonable time thereafter.

A handwritten signature in black ink that reads "Jonathan Mascolo".

Digitally signed under ESRA
by Jonathan Mascolo
on 10/24/2023 02:43 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**SIMULATOR
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Lake George Park Commission

Manufacturer: GUTH LABORATORIES

Model: 12V500

Serial Number: MP5081

Date: October 24, 2023

1. General appearance and condition of this device -----: OK

2. Check propeller rotation -----: OK

3. Check for air leaks -----: OK

4. Check for proper temperature -----: OK

**5. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**



**Division of Criminal
Justice Services**

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 10/24/2023 02:43 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

NIST Traceable
Calibration Report

REPORT NUMBER

1729415

Reference Number: 142177100

PO Number: SCARLUCCI0052423

NY State Div Of Criminal Justice Svcs
 80 South Swan Street
 Albany, NY 12210 United States

Manufacturer: Ertco
Model Number: 4400
Description: Temperature, Dig thermometer w/probe
Asset Number: 39808
Serial Number: 307443
Procedure: DS Eutechnics 4400-EP642

Calibration Date: 06/09/2023
Calibration Due Date: 06/09/2024
Condition As Found: In Tolerance
Condition As Left: In Tolerance, No adjustment

Remarks:

NIST-traceable calibration performed on the unit referenced above in accordance with customer requirements, published specifications and the lab's standard operating procedures. No adjustments were made to the unit.

Standards Used

Standard ID	Manufacturer	Model Number	Description	Cal Date	Due Date
CP144944	Fluke Corporation	1523	Temperature, Indicator	6/23/2022	6/30/2023
CP54452	Hart Scientific	5615-12	Temperature, Platinum RTD 12" Probe	12/19/2022	6/30/2023

Calibration Data

Function Tested	Nominal / Reference Value	Measured Value	OOT	Calibration Tolerance <i>g: = Guard Banding Applied</i>	TUR	EMU
Temperature Accuracy	-15.000 °C					
As Found & As Left	-15.000	-14.96		-15.050 to -14.950 °C	1.4:1	± 0.036 °C
	0.000 °C					
As Found & As Left	0.000	0.01		-0.030 to 0.030 °C	0.90:1	± 0.033 °C
	75.000 °C					
As Found & As Left	75.000	75.00		74.970 to 75.030 °C	0.83:1	± 0.036 °C
	125.000 °C					
As Found & As Left	125.000	125.02		124.950 to 125.050 °C	1.3:1	± 0.039 °C

Temperature: 24 °C
Humidity: 32 %RH
Rpt. No.: 1729415

Calibration Performed By:		Quality Reviewer:	
Arnel Bautista	Metrologist	Jeff Ziegler	6/9/2023
Name	Title	Name	Date

This report may not be reproduced, except in full, without written permission of Innocal. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCCL Z540-1-1994, 10CFR90, Appendix B, ISO 9002-94, and ISO 17025:2017. Conformance based on Simple Acceptance as a Decision Rule. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the remarks section.

*10-24-23

SERVICE AUTHORIZATION FORM
 Highway Safety Technology Unit
 80 South Swan Street
 Albany, NY 12210
 Ph: 518-485-7636 / 518-402-0689
 Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency <i>Lake George Park Commission</i>		Today's Date	
Address <i>75 Fort George Rd.</i>		City, State, ZIP <i>Lake George, N.Y. 12845</i>	
Contact Person <i>Lt. Joe Johns</i>		Contact Telephone <i>518-668-9347</i>	
Contact Email <i>jjohns@lGPC.state.ny.us</i>		Best Way to Reach You During Appointment Period	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT <u>GUTH</u> Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:			
Model Number <i>12V500</i>		Instrument Serial Number <i>MP5081</i>	
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -		Type of maintenance (check appropriate) <input type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)			
DMT/SIM: DMT <u>SIM</u> <u>W*</u> Keyboard Other: <u>communication cable</u> <u>Tubes</u> <u>*Power cable</u>		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.			
Signature <i>Lt. Joe Johns</i>		Date <i>10/20/23</i>	

DCJS USE ONLY

Export °C <u>34.01</u> <u>34.01</u> <u>34.01</u>	Digital °C <u>34.03</u> <u>34.03</u> <u>34.03</u>	Thermometer: <u>307443</u>	Calibration Due: <u>6/9/2024</u>
		Air Leak Test: <u>OK</u>	Propeller Rotation: <u>OK</u>
<i>*@ 10-24-2023</i>			