




STATE OF NEW YORK  
 DIVISION OF CRIMINAL JUSTICE SERVICES  
 4 TOWER PLACE  
 ALBANY, NEW YORK 12203-3702  
<http://criminaljustice.state.ny.us>

**CERTIFICATION PURSUANT TO CPLR 4518  
 OF RECORDS MAINTAINED IN THE  
 REGULAR COURSE OF BUSINESS**

I, John R. Digman, Assistant Director, Office of Public Safety, New York State Division of Criminal Justice Services, 4 Tower Place, Albany, New York 12203, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of rule 4518 of the Civil Practice Law and Rules, that the copies annexed hereto are exact photocopies of original records of the Office of Public Safety which are in my possession, custody and control.

I further certify that the original records of inspection /maintenance /calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 105208, performed on February 03, 2010 by Office of Public Safety employee MICHAEL J. HESS, of which the annexed are photocopies, were made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such records were made at the time such inspection /maintenance /calibration was performed or within a reasonable time thereafter, and further that it was the regular course of the Office of Public Safety's business to make such records at the time such inspection /maintenance /calibration was performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

State of New York  
 County of Albany  
 Sworn to before me  
 On February 05, 2010

  
 \_\_\_\_\_  
 John R. Digman  
 Assistant Director  
 Office of Public Safety

  
 \_\_\_\_\_  
 Notary Public, State of New York  
 Commission Expires 7/24/2010

Spencer Brooks  
 Notary Public, State of New York  
 No. 01BR6150085  
 Qualified in Albany County  
 Commission Expires July 24, 2010

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Name of Submitting Agency:** Warren County Sheriff's Office  
**Manufacturer:** NAT'L PATENT ANALYTICAL SYS  
**Model:** DATAMASTER DMT  
**Serial Number:** 105208  
**Date:** February 03, 2010

**Reference Solution Lot Number:** 09310  
**Reference Solution Standard:** 0.100 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.097 %	33.9 degrees centigrade
2	0.097 %	34.0 degrees centigrade
3	0.098 %	33.9 degrees centigrade
4	0.098 %	33.9 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 105208, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10(Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that the entries made in these records of inspection /maintenance /calibration were made at the time that such inspection /maintenance /calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

  
\_\_\_\_\_  
**MICHAEL J. HESS**  
Technician

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

**Manufacturer: NAT'L PATENT ANALYTICAL SYS**

**Model: DATAMASTER DMT**

**Serial Number: 105208**

**Date: February 03, 2010**

1. **General Appearance and condition of this device -----:**  
N/A
2. **Check controls and indicators -----: OK**
3. **Check breath flow system -----: OK**
4. **Check printer operation -----: OK**
5. **System cycles through all modes -----: OK**
6. **Record any repairs made to this instrument and list services provided:**  
RAN COMMUNICATION, DIAGNOSTIC AND SUPERVISOR TESTS,  
VERIFIED CALIBRATION, OK

  
\_\_\_\_\_  
**MICHAEL J. HESS**  
Technician  
Office of Public Safety