



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 105208, performed on July 07, 2020 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under
ESRA by Joshua Vinehout
on 7/27/2020 9:59 AM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office

Manufacturer: NAT'L PATENT ANALYTICAL SYS

Model: DATAMASTER DMT

Date: July 07, 2020

Serial Number: 105208

Reference Solution Lot Number: 19281

Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	34.00 degrees centigrade
2	0.100 %	33.99 degrees centigrade
3	0.100 %	34.00 degrees centigrade
4	0.100 %	34.00 degrees centigrade

Reference Solution Lot Number: 20020

Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.080 %	33.99 degrees centigrade
2	0.080 %	33.99 degrees centigrade
3	0.081 %	33.99 degrees centigrade
4	0.081 %	33.99 degrees centigrade

Reference Solution Lot Number: 19210

Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.181 %	33.99 degrees centigrade
2	0.181 %	34.00 degrees centigrade
3	0.181 %	33.99 degrees centigrade
4	0.181 %	33.99 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 105208, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under
ESRA by Joseph Iavarone
on 7/7/2020 1:52 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 7/8/2020 1:06 PM

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 105208
Date: July 07, 2020

1. General Appearance and condition of this device -----: OK

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: OK

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: **INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES, PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.**



Performed by:

Digitally signed under
ESRA by Joseph Iavarone
on 7/7/2020 1:52 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:

Digitally signed under
ESRA by Frank Sgarlata
on 7/8/2020 1:06 PM

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208





Date: 07/07/2020

OPERATOR NAME:
JOSEPH A IAVARONE
Permit #: 44106
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19281

Nominal = 0.100 g/210L

BLANK TEST	0.000	11:00
INTERNAL STANDARD	VERIFIED	11:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.100	11:00
BLANK TEST	0.000	11:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.100	11:01
BLANK TEST	0.000	11:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.100	11:02
BLANK TEST	0.000	11:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.100	11:04
BLANK TEST	0.000	11:04

Average = 0.1000
Std Dev = 0.0000



TECHNICIAN TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208





Date: 07/07/2020

OPERATOR NAME:
JOSEPH A IAVARONE
Permit #: 44106
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20020

Nominal = 0.080 g/210L

BLANK TEST	0.000	11:52
INTERNAL STANDARD	VERIFIED	11:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.080	11:53
BLANK TEST	0.000	11:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.080	11:54
BLANK TEST	0.000	11:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.081	11:55
BLANK TEST	0.000	11:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.081	11:56
BLANK TEST	0.000	11:57

Average = 0.0805
Std Dev = 0.0006



TECHNICIAN TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208





Date: 07/07/2020

OPERATOR NAME:
JOSEPH A IAVARONE
Permit #: 44106
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19210

Nominal = 0.180 g/210L

BLANK TEST	0.000	12:04
INTERNAL STANDARD	VERIFIED	12:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.181	12:04
BLANK TEST	0.000	12:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.181	12:06
BLANK TEST	0.000	12:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.181	12:07
BLANK TEST	0.000	12:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.181	12:08
BLANK TEST	0.000	12:09

Average = 0.1810
Std Dev = 0.0000



STATE OF NEW YORK
DIVISION OF CRIMINAL JUSTICE SERVICES


SERVICE AUTHORIZATION FORM

This authorization must accompany each piece of equipment when it is presented to the Equipment Repair Center for Service.

SECTION I - AGENCY INFORMATION

Name of Agency Warren County Sheriff's office		Today's Date 7/7/20
Address 1400 State Route 9		City, State, ZIP Lake George NY 12845
Contact Person Patrol officer T. Morse		Contact Telephone (518) 743-2500
Contact Telephone	Contact Facsimile	Contact Email tyler.morse@warrencountysheriffny.us

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer Pitmaster		Model Number DMT
Instrument Serial Number 105208	Antenna Serial Number (if applicable)	
Date of Malfunction	Type of maintenance (circle appropriate) <u>Calibration Check</u> Repair	
Brief Description of Malfunction		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units. DMT power cord, sim tubes, breath tube, keyboard simulator model 12V500 MP4824 sim power cord		
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein.		
Signature 		Date 7/7/2020

DCJS USE ONLY

inspected instrument, checked setup voltages, print and remote operation. performed calibration, adjustment procedure, ran subject technician and diagnostic tests, verified calibration.