



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 105208, performed on January 05, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance/calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under ESRA
by Joshua Vinehout
on 1/12/2022 1:23 PM

**Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety**

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: NAT'L PATENT ANALYTICAL SYS Model: DATAMASTER DMT
Serial Number: 105208
Date: January 05, 2022

Reference Solution Lot Number: 21390
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.102 %	34.00 degrees centigrade
2	0.101 %	34.00 degrees centigrade
3	0.102 %	34.00 degrees centigrade
4	0.102 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 105208, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under
ESRA by Jonathan Mascolo
on 1/5/2022 10:15 AM

Jonathan Mascolo
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45339

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 1/11/2022 9:40 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 105208
Date: January 05, 2022

- 1. General Appearance and condition of this device -----: N/A

- 2. Check controls and indicators -----: OK

- 3. Check breath flow system -----: OK

- 4. Check printer operation -----: N/A

- 5. System cycles through all modes -----: OK

- 6. Record any repairs made to this instrument and list services provided: **CHECKED REMOTE OPERATION, RAN DIAGNOSTIC AND TECHNICIAN TESTS, VERIFIED CALIBRATION.**



Performed by:

Digitally signed under
ESRA by Jonathan Mascolo
on 1/5/2022 10:15 AM

Jonathan Mascolo
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45339

Reviewed by:

Digitally signed under
ESRA by Frank Sgarlata
on 1/11/2022 9:40 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/05/2022

Nominal = 0.100 g/210L

BLANK TEST	0.000	09:49
INTERNAL STANDARD	VERIFIED	09:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.102	09:50
BLANK TEST	0.000	09:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.101	09:51
BLANK TEST	0.000	09:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.102	09:52
BLANK TEST	0.000	09:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.102	09:53
BLANK TEST	0.000	09:54

Average = 0.1017
Std Dev = 0.0005

Supervisor Record



Control Test Record

Serial Number: 105208

Test Date: 01/01/2022 13:12:48

Location: 366

WARREN CO. SHERIFF'S OFFICE

Supervisor Information

First Name: Charles B

Last Name: Geisler

Permit: 33757

Agency: WARREN CO. SO

Simulator Information

Lot: 21390 Nominal: 0.100

Test Results


Alcohol 1: 0.10 34.00

Alcohol 2: 0.10 34.00

Alcohol 3: 0.10 34.00

Solution Change:

Plot 1 of 3 <=>



Status

Status Code:

Date Time:

Message: