



STATE OF NEW YORK
DIVISION OF CRIMINAL JUSTICE SERVICES

Alfred E. Smith State Office Bldg.
80 South Swan Street
Albany, New York 12210
<http://criminaljustice.ny.gov>

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, John R. Digman, Special Programs Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the copies annexed hereto are exact photocopies of original records of the Office of Public Safety which are in my possession, custody and control.

I further certify that the original records of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 106606, performed on August 22, 2014 by Office of Public Safety employee Paul Hasbrouck, of which the annexed are photocopies, were made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such records were made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it was the regular course of the Office of Public Safety's business to make such records at the time such inspection/ maintenance/calibration was performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

State of New York
County of Albany
Sworn to before me
On August 27, 2014


John R. Digman
Special Programs Supervisor
Office of Public Safety


Notary Public, State of New York
Commission Expires 10/30/17

KEVIN J. BOEHME
NOTARY PUBLIC, STATE OF NEW YORK
QUALIFIED IN COLUMBIA COUNTY
01BO4958192
COMMISSION EXPIRES OCTOBER 30, 20 17

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office

Manufacturer: NAT'L PATENT ANALYTICAL SYS

Model: DATAMASTER DMT

Date: August 22, 2014

Serial Number: 106606

Reference Solution Lot Number: 14160

Reference Solution Standard: 0.10% at 34.0degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.102 %	34.00 degrees centigrade
2	0.102 %	34.01 degrees centigrade
3	0.102 %	34.01 degrees centigrade
4	0.102 %	34.00 degrees centigrade

Reference Solution Lot Number: 13251

Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.081 %	34.02 degrees centigrade
2	0.081 %	34.00 degrees centigrade
3	0.080 %	34.01 degrees centigrade
4	0.081 %	34.00 degrees centigrade

Reference Solution Lot Number: 14070

Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.182 %	34.00 degrees centigrade
2	0.182 %	34.01 degrees centigrade
3	0.182 %	33.99 degrees centigrade
4	0.181 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 106606, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.


I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:

Reviewed by:



Paul Hasbrouck
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 30738



Lisa A. Lee
Highway Safety Equipment Technical Supervisor
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 106606
Date: August 22, 2014


1. General Appearance and condition of this device -----: **OK, LAB MAINT.**
2. Check controls and indicators -----: **OK**
3. Check breath flow system -----: **OK**
4. Check printer operation -----: **OK**
5. System cycles through all modes -----: **OK**
6. Record any repairs made to this instrument and list services provided: **RUN DIAGNOSTIC, TECHNICIAN, SUPERVISOR, SUBJECT, AND COMMUNICATION TEST, VERIFIED CALIBRATION, OK**

Performed by:



Paul Hasbrouck
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 30738

Reviewed by:



Lisa A. Lee
Highway Safety Equipment Technical Supervisor
Breath Analysis Technical Supervisor