



STATE OF NEW YORK
DIVISION OF CRIMINAL JUSTICE SERVICES
4 TOWER PLACE
ALBANY, NEW YORK 12203-3702
<http://criminaljustice.state.ny.us>

**CERTIFICATION PURSUANT TO CPLR 4518
OF RECORDS MAINTAINED IN THE
REGULAR COURSE OF BUSINESS**

I, Todd R. Murray, Technical Training Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 4 Tower Place, Albany, New York 12203, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of rule 4518 of the Civil Practice Law and Rules, that the copies annexed hereto are exact photocopies of original records of the Office of Public Safety which are in my possession, custody and control.

I further certify that the original records of inspection /maintenance /calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 144506, performed on March 02, 2011 by Office of Public Safety employee MICHAEL J. HESS, of which the annexed are photocopies, were made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such records were made at the time such inspection /maintenance /calibration was performed or within a reasonable time thereafter, and further that it was the regular course of the Office of Public Safety's business to make such records at the time such inspection /maintenance /calibration was performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

Todd R. Murray
Technical Training Supervisor
Office of Public Safety

State of New York
County of Albany
Sworn to before me
On March 04, 2011

Notary Public, State of New York
Commission Expires 7/24/2014

Spencer Brooks
Notary Public, State of New York
No. 01BR6150085
Qualified in Albany County
Commission Expires July 24, 2014

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 144506
Date: March 02, 2011

Reference Solution Lot Number: 10270
Reference Solution Standard: 0.100 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.095 %	34.0 degrees centigrade
2	0.096 %	33.9 degrees centigrade
3	0.096 %	33.9 degrees centigrade
4	0.096 %	34.0 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 144506, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10(Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that the entries made in these records of inspection /maintenance /calibration were made at the time that such inspection /maintenance /calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.



MICHAEL J. HESS
Technician

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Manufacturer: NAT'L PATENT ANALYTICAL SYS

Model: DATAMASTER DMT

Serial Number: 144506

Date: March 02, 2011

1. **General Appearance and condition of this device -----:**
N/A
2. **Check controls and indicators -----: OK**
3. **Check breath flow system -----: OK**
4. **Check printer operation -----: OK**
5. **System cycles through all modes -----: OK**
6. **Record any repairs made to this instrument and list services provided:**
RAN COMMUNICATION, DIAGNOSTIC AND SUPERVISOR TESTS,
VERIFIED CALIBRATION, OK



MICHAEL J. HESS
Technician
Office of Public Safety