




STATE OF NEW YORK  
 DIVISION OF CRIMINAL JUSTICE SERVICES  
 Alfred E. Smith State Office Bldg.  
 80 South Swan Street  
 Albany, New York 12210  
<http://criminaljustice.ny.gov>

**CERTIFICATION PURSUANT TO CPLR 4518  
 OF RECORDS MAINTAINED IN THE  
 REGULAR COURSE OF BUSINESS**

I, John R. Digman, Special Programs Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of rule 4518 of the Civil Practice Law and Rules, that the copies annexed hereto are exact photocopies of original records of the Office of Public Safety which are in my possession, custody and control.

I further certify that the original records of inspection /maintenance /calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 144506, performed on August 19, 2013 by Office of Public Safety employee Frank Sgarlata, of which the annexed are photocopies, were made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such records were made at the time such inspection /maintenance /calibration was performed or within a reasonable time thereafter, and further that it was the regular course of the Office of Public Safety's business to make such records at the time such inspection /maintenance /calibration was performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

State of New York  
 County of Albany  
 Sworn to before me  
 On August 21, 2013

  
 \_\_\_\_\_  
 John R. Digman  
 Special Programs Supervisor  
 Office of Public Safety

  
 \_\_\_\_\_  
 Notary Public, State of New York  
 Commission Expires \_\_\_\_\_

KEVIN J. BOEHME  
 NOTARY PUBLIC, STATE OF NEW YORK  
 QUALIFIED IN COLUMBIA COUNTY  
 # 01BO4958192  
 COMMISSION EXPIRES OCTOBER 30, 2013

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office  
Manufacturer: NAT'L PATENT ANALYTICAL SYS  
Model: DATAMASTER DMT                      Serial Number: 144506  
Date: August 19, 2013

Reference Solution Lot Number: 12171  
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.098 %	34.02 degrees centigrade
2	0.097 %	34.01 degrees centigrade
3	0.098 %	34.01 degrees centigrade
4	0.097 %	34.01 degrees centigrade

Reference Solution Lot Number: 12181  
Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.079 %	34.00 degrees centigrade
2	0.081 %	34.01 degrees centigrade
3	0.080 %	34.01 degrees centigrade
4	0.080 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 144506, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10(Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that the entries made in these records of inspection /maintenance /calibration were made at the time that such inspection /maintenance /calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

  
Frank Sgarlata  
Highway Safety Equipment Technician  
Office of Public Safety

**BREATH TEST INSTRUMENT  
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**


**Manufacturer: NAT'L PATENT ANALYTICAL SYS**

**Model: DATAMASTER DMT**

**Serial Number: 144506**

**Date: August 19, 2013**

1. General Appearance and condition of this device -----:  
OK.
2. Check controls and indicators -----: OK
3. Check breath flow indicator -----: OK
4. Check printer operation -----: OK
5. System cycles through all modes -----: OK
6. Record any repairs made to this instrument and list services provided:  
CALIBRATED INSTRUMENT. RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC  
TESTS. CHECKED SETUP VOLTAGES, PRINT AND REMOTE OPERATION.  
VERIFIED CALIBRATION.



---

**Frank Sgarlata**  
**Highway Safety Equipment Technician**  
**Office of Public Safety**