



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NATL PATENT ANALYTICAL SYSTEM, DATAMASTER DMT breath test instrument, serial number 144506, performed on January 06, 2023 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 01/12/2023 08:50 AM

**Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety**

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 144506
Date: January 06, 2023


Reference Solution Lot Number : 22390
Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.098 %	34.00 degrees centigrade
2	0.099 %	34.00 degrees centigrade
3	0.099 %	34.00 degrees centigrade
4	0.100 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NATL PATENT ANALYTICAL SYSTEM breath test instrument model DATAMASTER DMT, serial number 144506, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under ESRA
by Joseph Iavarone
on 01/06/2023 12:42 PM

Reviewed by:



Digitally signed under ESRA
by Frank Sgarlata
on 01/06/2023 01:24 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit #44106

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Warren County Sheriff's Office
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 144506
Date: January 06, 2023

- 1. General appearance and condition of this device -----: N/A

- 2. Check controls and indicators -----: OK

- 3. Check breath flow system -----: OK

- 4. Check printer operation -----: N/A

- 5. System cycles through all modes -----: OK

- 6. Record any repairs made to this instrument and list services provided : CHECKED REMOTE OPERATION, RAN DIAGNOSTIC AND TECHNICIAN TESTS, VERIFIED CALIBRATION.



Performed by:

Digitally signed under ESRA
by Joseph Iavarone
on 01/06/2023 12:42 PM

Reviewed by:

Digitally signed under ESRA
by Frank Sgarlata
on 01/06/2023 01:24 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit #44106

Frank Sgarlata
Breath Analysis Technical Supervisor

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/06/2023

Nominal = 0.100 g/210L

BLANK TEST	0.000	10:09
INTERNAL STANDARD	VERIFIED	10:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.098	10:09
BLANK TEST	0.000	10:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	10:10
BLANK TEST	0.000	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	10:11
BLANK TEST	0.000	10:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.100	10:13
BLANK TEST	0.000	10:14

Average = 0.0990
Std Dev = 0.0008

Control Test Record

Serial Number: 144506

Test Date: 01/02/2023 00:47:23

Location: 366

WARREN CO. SHERIFF'S OFFICE

Supervisor Information

First Name: Michael A

Last Name: Campbell

Permit: 42422

Agency: WARREN CO. SO

Simulator Information

Lot: 22390 Nominal: 0.100

Test Results


Alcohol 1: 0.09 34.00

Alcohol 2: 0.10 34.00

Alcohol 3: 0.10 34.00

Solution Change:

Plot 1 of 3 <=>



Status

Status Code:

Date Time:

Message: