

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 01/11/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 07250

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:01
INTERNAL STANDARD	VERIFIED	19:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	19:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	19:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	19:03
BLANK TEST	0.00	19:04



SGT *Ralph Bartlett* #162