

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/26/2007

Operator Name: C. S LAMOUREE
Permit #: 23039
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:41
INTERNAL STANDARD	VERIFIED	19:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	INTERFERENCE	19:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	INTERFERENCE	19:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	INTERFERENCE	19:44
BLANK TEST	0.00	19:45

