

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/28/2007

Operator Name: C S LAMOUREE
Permit #: 23039
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07180

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 20:31 |
| INTERNAL STANDARD | VERIFIED | 20:31 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 20:31 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 20:32 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 20:33 |
| BLANK TEST | 0.00 | 20:34 |

