

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 12/11/2007

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN COUNTY SHERIFF'S OFFICE  
Reference Standard #: 07250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:47
INTERNAL STANDARD	VERIFIED	00:48
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:48
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:49
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51

SGT *Ralph Bartlett*  
12/11/07

