

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/15/2007

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN COUNTY SHERIFF'S OFFICE
Reference Standard #: 07250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:20
INTERNAL STANDARD	VERIFIED	02:20
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:20
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:23



SGT *Ralph Bartlett* 12/15/07