

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/31/2007

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN COUNTY SHERIFF'S OFFICE
Reference Standard #: 07250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:49
INTERNAL STANDARD	VERIFIED	03:49
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:49
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:50
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:52



Sgt Ralph J Bartlett

12/31/07