

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 01/14/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 07250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:41
INTERNAL STANDARD	VERIFIED	03:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:43
REFERENCE STANDARD TEMP	34.0c	
BLANK TEST	0.00	03:44



505 Ralph Bartlett # 162