

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/31/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07300

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:11
INTERNAL STANDARD	VERIFIED	05:11
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:11
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:12
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:13
BLANK TEST	0.00	05:14



SG Ralph Bartlett 1/31/08