

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 02/06/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 07300

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:18
INTERNAL STANDARD	VERIFIED	02:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:19
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:20
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:21



SGT Ralph Bartlett SHIRO # 162 2/6/08