

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/13/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07300

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:04
INTERNAL STANDARD	VERIFIED	19:05
REFERENCE STANDARD TEMP	33.9c	
REFERENCE STANDARD	0.09	19:05
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	19:06
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	19:07
BLANK TEST	0.00	19:08



SGT Ralph J Bartlett # 162

2/13/08

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/13/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07300

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:23
INTERNAL STANDARD	ERROR	18:23

INTERNAL STANDARD ERROR