

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/29/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07300

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:18
INTERNAL STANDARD	VERIFIED	04:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:19
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:21
BLANK TEST	0.00	04:21



SGT *Ralph Bartlett* 2/29/08