

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/09/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07300

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:21
INTERNAL STANDARD	VERIFIED	06:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	INTERFERENCE	06:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	06:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	06:23
BLANK TEST	0.00	06:24



SGT *Rushy Bartlett* #162