

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/23/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 07300

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:13
INTERNAL STANDARD	VERIFIED	18:13
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:15
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:16
BLANK TEST	0.00	18:16



SGT Ralph Bartlett #162