

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 04/01/2008

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 08010

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:17
INTERNAL STANDARD	VERIFIED	03:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD INTERFERENCE		03:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:19
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21



SGT *Ralph J Bartlett* WESO # 162