

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 04/07/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08010

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:42
INTERNAL STANDARD	VERIFIED	03:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:44
REFERENCE STANDARD TEMP	34.0c	
BLANK TEST	0.00	03:45



SOS Ralph Bartlett #162