

# SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 04/15/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08010

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 06:12 |
| INTERNAL STANDARD       | VERIFIED | 06:13 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 06:13 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 06:14 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 06:15 |
| BLANK TEST              | 0.00     | 06:16 |



SGT Ralph J Bartlett

4/15/08