

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 04/21/2008

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 08010

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:01
INTERNAL STANDARD	VERIFIED	04:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD INTERFERENCE		04:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04



SG Sheriff Bartlett 4/21/08