

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/03/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08010

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:48
INTERNAL STANDARD	VERIFIED	00:48
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:49
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:50
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:51



SGT Ralph Bartlett Shield 162