

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/12/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08010

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:04
INTERNAL STANDARD	VERIFIED	19:04
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	19:04
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	19:05
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	19:06
BLANK TEST	0.00	19:07



SGT Ralph Bartlett #162