

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/19/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08010

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:30
INTERNAL STANDARD	VERIFIED	03:30
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:30
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:31
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33



SGT Ralph J Bartlett # 162 MAY 19, 2008.