

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/26/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08010

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:21
INTERNAL STANDARD	VERIFIED	18:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	18:23
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:24
BLANK TEST	0.00	18:24



SGS Ralph Bartlett 5/26/08