

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 06/05/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08070

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:04
INTERNAL STANDARD	VERIFIED	23:04
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	INTERFERENCE	23:04
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	23:05
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	23:06
BLANK TEST	0.00	23:07



SOS *Ralph Bartlett* Sheriff 162 6/5/08