

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/13/2008

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08070

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:42
INTERNAL STANDARD	VERIFIED	16:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	16:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	16:44
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	16:45
BLANK TEST	0.00	16:46



J. Lopez