

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 06/19/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08070

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:04
INTERNAL STANDARD	VERIFIED	21:04
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	INTERFERENCE	21:05
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	21:06
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	21:07
BLANK TEST	0.00	21:07



505 Dept Sheriff 6/19/08