

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 07/02/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08070

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 20:02 |
| INTERNAL STANDARD       | VERIFIED | 20:02 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.10     | 20:03 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.10     | 20:04 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.10     | 20:05 |
| BLANK TEST              | 0.00     | 20:06 |



SGS Ralph Bartlett 7/2/08

WCSO #162