

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/01/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08160

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 04:08 |
| INTERNAL STANDARD | VERIFIED | 04:08 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 04:08 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.10 | 04:09 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.10 | 04:10 |
| BLANK TEST | 0.00 | 04:11 |

