

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/07/2008

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08160

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 09:59 |
| INTERNAL STANDARD | VERIFIED | 09:59 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.10 | 09:59 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.10 | 10:00 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.10 | 10:01 |
| REFERENCE STANDARD TEMP | 34.0c | |
| BLANK TEST | 0.00 | 10:02 |

Charles B. Geisler #248

