

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/13/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08160

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:18
INTERNAL STANDARD	VERIFIED	18:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	18:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	18:19
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	18:20
BLANK TEST	0.00	18:21



SGS RALPH BARTLETT #162