

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/22/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08160

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:58
INTERNAL STANDARD	VERIFIED	18:58
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:58
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:59
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01



SOB *Ralph Bartlett*

SHERIFF # 112

8/22/08