

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 08/29/2008

Operator Name: CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08160

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:00
INTERNAL STANDARD	VERIFIED	07:00
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	07:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	07:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	07:03
BLANK TEST	0.00	07:03

*po C.B. Geisler #248*

*8/29/08*

