

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 09/08/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08160

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:37
INTERNAL STANDARD	VERIFIED	03:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:38
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:39
BLANK TEST	0.00	03:40



505 *Ralph J Bartlett* 9/8/08 SHIELD # 162