

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/16/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08160

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:51
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:51
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:52
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54



SGT Ralph J Bartlett # 162